



ALFATEST - 40LP

Water Meters; Loss of Pressure Testing Equipment

TRADEMARK
ALFA

Max. pressure-loss
 $\Delta P 63 = 0,63 \text{ bar}$

Nominal Diameter
Dn15~Dn40

ORIGIN
TURKEY

Relative Humidity
45~75%,rh

Air Temperature
 $20 \pm 5^\circ \text{C}$

Basic Technical Details

- Where it has been established that the pressure loss of the meter will follow the square law.
- The pressure loss shall be tested at **Q₃ only**.
- When it is suspected that, a pressure loss peak occurs below Q₃, the pressure loss shall be determined between Q₁ - Q₃, starting at Q₁ and increasing the flow rate by max. 0,1 x Q₃.

After Q₃ is reached, the flow rate shall be decreased by max. 0,1 x Q₃.

(ISO 4064-1:2014 | OIML R 49-1:2013, 6.5)



ISO 4064-1:2014

alfawatermeter@gmail.com

Macun Mah. Anadolu Blv. ATB İş Merkezi K Blok
No.278 Yenimahalle / Ankara - Türkiye



www.alfawatermeter.com